

Research-Grade VPP Materials for Scaffold and Hydrogel Studies

Porous, rigid and responsive materials requiring cell-system and application-specific evaluation

Dr. Juan Segurola, founder and Scientific Director · 3Dresyns by Resyner Technologies S.L. · General Etxague 15B, 20003 Donostia / San Sebastián, Gipuzkoa, Spain · info@3dresyns.com · www.3Dresyns.com

Highlights

- The portfolio is presented as research-grade materials for controlled geometry studies.
- Biological evidence is identified by exact formulation, processing, endpoint, cell system and report.
- Mechanical rigidity, porosity and biological response are separate attributes.
- Assay interference, adsorption, leachables, autofluorescence and sterilisation must be evaluated where relevant.
- The products are not pre-validated implants, therapies or finished medical devices.

Abstract

The portfolio comprises research-grade photopolymer materials for porous scaffolds, rigid lattices and responsive hydrogel structures. Controlled geometry can support experimental studies, but biological suitability depends on exact formulation, printing, cleaning, post-curing, sterilisation, surface condition and biological model.

Where product-specific biological data exist, the tested material, conditions and endpoint are identified rather than applying biocompatible categorically across a family. Cell viability, proliferation, differentiation, adsorption and extractables are distinct outcomes.

The products are not pre-validated medical devices, implants or regenerative-medicine therapies. Translation beyond research requires finished-device risk management, manufacturing validation, biological evaluation and the applicable regulatory pathway.


Keywords research scaffold · hydrogel · VPP · cell culture · organ-on-chip · biological evaluation

Source use. Product and biological reports are controlled in [1]; scaffold and bioprinting context in [2–3,6]; biological and quality-system scope in [4–5].

1. Research-material classification

Porous scaffold grades, a rigid lattice grade and hydrogel/responsive grades are treated as separate research-material classes. Portfolio membership does not establish a shared biological response or clinical indication.

- Record exact formulation and controlled document.
- Separate geometry and mechanics from biological performance.
- Keep research-grade visible in every application map.



Release requires traceable evidence at every gate; a catalogue label cannot bypass the chain.

Figure 1. WP-20 evidence gates from controlled material description to final acceptance.

2. Biological-study record

A claim record identifies specimen geometry, printing and post-cure, cleaning, sterilisation, extraction or direct contact, endpoint, cell line or biological model, controls, laboratory report and limitation.

- Do not infer viability, proliferation or differentiation.
- Check adsorption, leachables and autofluorescence.
- Assess whether the material changes assay readout or device function.

3. Translation boundary

Tissue-engineering, organ-on-chip and regenerative-medicine descriptions are research implementation candidates. ISO 13485 is a quality-management-system standard and does not test scaffold biological performance. Finished-device claims require a separate legal and evidence pathway.

4. Minimum controlled evidence before release

Control field	Required record	Release interpretation
Material	Exact formulation and processed specimen	Catalogue description only until evidence is linked
Study	Geometry, sterilisation and biological model	Control process variables and change triggers
Endpoints	Viability, function, adsorption and leachables	Report method, result, uncertainty and limits
Translation	Risk, manufacturing, biology and conformity	Apply final acceptance and responsibility boundary



Description ≠ measured result ≠ certification or finished-device approval

Figure 2. WP-20 separates catalogue positioning, measured evidence and final application decisions.

KEY INSIGHT

Precise product or application language is retained only when the exact formulation, method, processed specimen and controlled evidence can be traced.

5. Limitations

- Research-grade does not mean implantable, therapeutic or clinically validated.
- Biological results are test-system and processing-specific.
- Mechanical rigidity does not establish biological suitability.
- A Bio designation is not a biological endpoint.
- Translation requires finished-device evidence and legal-manufacturer control.

6. Conclusions

- The portfolio is presented as research-grade materials for controlled geometry studies.
- Biological evidence is identified by exact formulation, processing, endpoint, cell system and report.
- The products are not pre-validated implants, therapies or finished medical devices.
- This corrected candidate remains local and on HOLD until series-wide scientific, metadata and visual release checks are complete.

Revision note

Version 5.3 removes categorical scaffold biocompatibility and clinical implications, adds exact test- and cell-system evidence requirements, keeps research-grade boundaries visible and updates ISO 10993-1 to 2025.

Acknowledgements. The author acknowledges the 3Dresyns technical team and users whose manufacturing experience informed this correction.

Disclosure. Dr. Juan Segurola is founder and Scientific Director of 3Dresyns by Resyner Technologies S.L., the supplier of the materials discussed.

References

- [1] 3Dresyns. Current scaffold/hydrogel product-specific TDS, IFU and controlled biological reports.
- [2] Hollister, S. J. (2005). Porous scaffold design for tissue engineering. *Nature Materials*, 4, 518–524. <https://doi.org/10.1038/nmat1421>
- [3] Murphy, S. V., & Atala, A. (2014). 3D bioprinting of tissues and organs. *Nature Biotechnology*, 32, 773–785. <https://doi.org/10.1038/nbt.2958>
- [4] ISO 10993-1:2025. Biological evaluation of medical devices—Part 1.
- [5] ISO 13485:2016. Medical devices—Quality management systems—Requirements for regulatory purposes.
- [6] Bagheri, A., & Jin, J. (2019). Photopolymerization in 3D printing. <https://doi.org/10.1021/acsapm.8b00165>